



PRUVENTOR™ FOAM HEEL OFF-LOADING DEVICE

	NOT MADE WITH NATURAL RUBBER LATEX
	SINGLE PATIENT USE ONLY
	NON-STERILE
RX ONLY	Federal U.S.A. law restricts this device to sale or use by or on the order of a physician (or properly licensed practitioner).

IMPORTANT INFORMATION

Please read all warnings and instructions before use. Correct application is essential for proper product function and to reduce the risk of injury or re-injury inherent with the use of any brace.

INTENDED USE

The DeRoyal PRUventor™ Foam Heel Off-loading Device is intended to be prescribed by a properly licensed practitioner to help prevent heel decubitus ulcers, foot drop and foot rotation.

CONTRAINDICATIONS

The PRUventor™ Foam Heel Off-loading Device is for non-ambulatory patients only and should be removed before ambulation. Do not allow your patients to stand or walk while wearing the boot.

WARNINGS

- Inspect device for damaged or missing components prior to patient application.
- Consult physician or properly licensed practitioner immediately if you experience sensation changes, unusual reactions, swelling or increased pain while using this product.
- Discontinue use and consult physician or properly licensed practitioner if device does not off-load the heel, or help prevent foot drop or foot rotation (if used with a wedge.)
- If device is used with a sequential/intermittent compression device, be sure tubing does not interfere with off-loading of the patient's foot, ankle or calf.
- Discontinue use of the device and consult physician or properly licensed practitioner if it becomes damaged in any way, limiting its ability to perform as intended.

CAUTIONS

- This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which it is intended. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care, as well as the patient.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.
- Be sure to follow your facility's policies and guidelines for frequency of patient monitoring.

INSTRUCTIONS FOR USE

- Release the hook tabs and loosen the straps.
- Remove the foam block that is inside the boot.
- Position the patient's foot inside the boot, with the heel suspended over the heel opening.
- Feed the straps through the D-ring buckles and secure hook closures.
- Visually check to ensure the straps do not restrict the patient's circulation.
- To help prevent hip/foot rotation, remove the adhesive backing from the foam block insert and place on the medial or lateral side of the boot.
- The patient's skin should be assessed frequently and the boot should be removed and placed back on the patient as specified by hospital policies and procedures. **NOTE:** Perform periodic skin assessment as specified by hospital policies and procedures.

SIZING

Universal



PRUVENTOR™ FOAM HEEL OFF-LOADING DEVICE

	NOT MADE WITH NATURAL RUBBER LATEX
	SINGLE PATIENT USE ONLY
	NON-STERILE
RX ONLY	Federal U.S.A. law restricts this device to sale or use by or on the order of a physician (or properly licensed practitioner).

IMPORTANT INFORMATION

Please read all warnings and instructions before use. Correct application is essential for proper product function and to reduce the risk of injury or re-injury inherent with the use of any brace.

INTENDED USE

The DeRoyal PRUventor™ Foam Heel Off-loading Device is intended to be prescribed by a properly licensed practitioner to help prevent heel decubitus ulcers, foot drop and foot rotation.

CONTRAINDICATIONS

The PRUventor™ Foam Heel Off-loading Device is for non-ambulatory patients only and should be removed before ambulation. Do not allow your patients to stand or walk while wearing the boot.

WARNINGS

- Inspect device for damaged or missing components prior to patient application.
- Consult physician or properly licensed practitioner immediately if you experience sensation changes, unusual reactions, swelling or increased pain while using this product.
- Discontinue use and consult physician or properly licensed practitioner if device does not off-load the heel, or help prevent foot drop or foot rotation (if used with a wedge.)
- If device is used with a sequential/intermittent compression device, be sure tubing does not interfere with off-loading of the patient's foot, ankle or calf.
- Discontinue use of the device and consult physician or properly licensed practitioner if it becomes damaged in any way, limiting its ability to perform as intended.

CAUTIONS

- This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which it is intended. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care, as well as the patient.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.
- Be sure to follow your facility's policies and guidelines for frequency of patient monitoring.

INSTRUCTIONS FOR USE

- Release the hook tabs and loosen the straps.
- Remove the foam block that is inside the boot.
- Position the patient's foot inside the boot, with the heel suspended over the heel opening.
- Feed the straps through the D-ring buckles and secure hook closures.
- Visually check to ensure the straps do not restrict the patient's circulation.
- To help prevent hip/foot rotation, remove the adhesive backing from the foam block insert and place on the medial or lateral side of the boot.
- The patient's skin should be assessed frequently and the boot should be removed and placed back on the patient as specified by hospital policies and procedures. **NOTE:** Perform periodic skin assessment as specified by hospital policies and procedures.

SIZING

Universal



CLEANING AND/OR MAINTENANCE

To clean, hand wash with a mild fabric detergent in warm water. Do not bleach. Do not iron. Lay flat and open and allow to air dry. Inspect the product after washing to ensure it is intact. If the product is damaged, do not use it and replace it with a new one.

WARRANTY

DeRoyal products are warranted for one hundred twenty (120) days from the date of shipment from DeRoyal as to product quality and workmanship. **DEROYAL'S WRITTEN WARRANTIES ARE GIVEN IN LIEU OF ANY IMPLIED WARRANTIES, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.**



Manufacturer
DeRoyal Industries, Inc.
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#0-1965G Revised 1/17

©2017 DeRoyal Industries, Inc.
All Rights Reserved. PRUventor, DeRoyal, the DeRoyal logo, and Improving Care. Improving Business. are trademarks or registered trademarks of DeRoyal Industries, Inc.



CLEANING AND/OR MAINTENANCE

To clean, hand wash with a mild fabric detergent in warm water. Do not bleach. Do not iron. Lay flat and open and allow to air dry. Inspect the product after washing to ensure it is intact. If the product is damaged, do not use it and replace it with a new one.

WARRANTY

DeRoyal products are warranted for one hundred twenty (120) days from the date of shipment from DeRoyal as to product quality and workmanship. **DEROYAL'S WRITTEN WARRANTIES ARE GIVEN IN LIEU OF ANY IMPLIED WARRANTIES, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.**



Manufacturer
DeRoyal Industries, Inc.
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#0-1965G Revised 1/17

©2017 DeRoyal Industries, Inc.
All Rights Reserved. PRUventor, DeRoyal, the DeRoyal logo, and Improving Care. Improving Business. are trademarks or registered trademarks of DeRoyal Industries, Inc.