Head Office 11-13 Scoresby Road 3153 Australia Bayswater Victoria Ph +61 3 9720 8840

Fx +61 3 9720 8860

www.fairmontmedical.com.au

EC REP

www.fairmontmedical.com Ph +44 (0) 1245 206800 28 Baddow Road Chelmsford Elizabeth House Fairmont Medical Products Ltd

Fairmont Medical Products Pty Ltd is certified to ISO13485:2003

Issued by E. NITEROS 26/06/2018 PM014 Version 1.5

Product specifications subject to change without notice. Pictures shown may vary from actual product.
© 2018 Fairmont Medical Products Pty Ltd all rights reserved.

-Fairmont

MEDICAL



For more information on this product please scan QR code

Instructions For Use

Disposable Urology Set and Disposable Irrigation Set























Caution - Read all instructions prior to use

The following instructions relate to the following products: DUS/DIT Range

Indications / Intended Use

well as associated anatomy such as urethra and the working unit connecting a urology working unit / inspection sheath or allow irrigation primarily of the bladder as The Disposable Urology Set (DUS) and Disposable Irrigation Set (DIT) range is indicated for

DUS4005 / DUS4015 / DUS4215 / DIT5015 / DIT5028

- Peel the pouch open using standard aseptic techniques and pass on inner bag to sterile
- Sterile nurse to remove from inner bag, uncoil sterile device and pass bag spike to circulating nurse again using standard aseptic technique.
- Circulating nurse to engage roller clamp to firmly restrict flow.
- ω 4. τ Circulating nurse to remove bag spike cover whilst holding the bag spike handle
- Whilst holding the bag spike handle push bag spike firmly into prepared irrigant bag outlet using universal aseptic technique. The bag spike must be inserted firmly enough as to not become dislodged during procedure.
- 7.6 Sterile nurse to connect flexible tubing (distal end) to urology working unit Luer port
- Circulating nurse to disengage roller clamp to allow flow to urology working unit. Flow may be visually sighted by viewing the Flow Chamber in DUS products.
- At the conclusion of the procedure, engage roller-clamp and remove from Urology working

*In most applications the roller clamp is set to the fully open position. However the roller clamp can be used to restrict flow to the desired clinical application.

DUS4008 / DUS4018 / DUS4021 / DUS4218 / DUS4318 / DUS4518

- Peel DUS pouch open using standard aseptic techniques and pass on inner bag to sterile
- 2 Sterile nurse to remove DUS from inner bag, uncoil sterile device and pass bag spike to circulating nurse again using standard aseptic technique.
- Circulating nurse to engage G clamps firmly with an audible click.
- ω 4. Circulating nurse to remove bag spike covers(s) one at a time and discard whilst holding the bag spike handle.
- 5 Whilst holding the bag spike handle, push bag spike(s) firmly into prepared irrigant bag outlet using universal aseptic technique. The spike(s) must be inserted firmly enough as to not become dislodged during procedure.
- 6. Sterile nurse to connect flexible tubing or Luer connector (DUS4021) to urology working unit Luer port.
- Circulating nurse to unclip one G clamp to allow flow to urology working unit. Flow may be visually sighted by viewing the Flow Chamber.
- At the conclusion of the procedure, engage G clamp(s) and remove from Urology working

In most applications the roller clamp is set to the fully open position. However the roller clamp can be used to restrict flow to the desired clinical application.

E

Foley catheter Optional connection to

post-surgery

- Circulating nurse to apply G clamp(s).
- Sterile nurse to remove from urology working firmly holding the universal connector. unit, pull firmly on the flexible tube whilst
- Firmly push universal connector into Foley

fluid bag Optional transfer / replacement of irrigation

swapped intraoperatively If required, additional bags of irrigant may be

- desired line. Circulating nurse to apply G clamp on
- and pull to remove the bag spike utilizing Firmly hold bag spike and gently twist universal aseptic technique.
- Grasp the bag spike handle irrigant bag
- required Release G clamp on irrigation line when

Precautions

healthcare personnel. Not to be used for any other This device may only be used by suitably trained Do not use if package is compromised/damaged

or removing DUS/DIT from irrigant bag. IFU pertains to Sterile devices sold by Fairmont please refer to 3rd party manufacturer for sterility Medical. If the device has code "XXXX NS" This device is sterile (ETO). Please note, this Single use only, do not re-sterilize. Always grip bag spike handle firmly when inserting

Materials

irrigation tubing sets made primarily of The DUS/DIT range is sterile, double-wrapped Phthalate-free PVC.

All components and packaging is Latex free The flexible extension tube is Latex free

Flexible Tubing

